







SQUARE ABUTMENTS & INTERFACES

DIRECT TO IMPLANT									
CONNECTION	Square abutments			Screw included	MAXIMUM ANGULATION screw included	TORQUE RECOMMENDATION screw included	ANGLED SCREW and maximum angulation	TORQUE RECOMMENDATION angled screws (Ncm)	PROSTHETIC Ø (mm)
Interna 3.0 		INPSP30	(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4
		INPSP31	(2)(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4
		INPSP32	(2)(4)	INTTUHCR	14°	0°: 35 ≤14°: 20	-	-	4
		INPSP33	(2)(4)	INTTUHCR	7°	0°: 35 ≤7°: 20	-	-	4
		INPSP3R0	(1)(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	3.8
Interna Narrow (4) 		INPSPE0	(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4
		INPSPE1	(2)(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4
		INPSPE2	(2)(4)	INTTUHCR	7°	0°: 35 ≤7°: 20	-	-	4
		INPSPE3	(2)(4)	INTTUHCR	5°	0°: 35 ≤5°: 20	-	-	4
		INSPER0	(1)(4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	3.8
		INSPER1	(1)(2) (4)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	3.8
		INSPER2	(1)(2) (4)	INTTUHCR	7°	0°: 35 ≤7°: 20	-	-	3.8
		INSPER3	(1)(2) (4)	INTTUHCR	5°	0°: 35 ≤5°: 20	-	-	3.8
		INPSPEX0	(3)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4.4
		INPSPEX1	(2)(3)	INTTUHCR	15°	0°: 35 ≤15°: 20	-	-	4.4
		INPSPEX2	(2)(3)	INTTUHCR	10°	0°: 35 ≤10°: 20	-	-	4.4
		INPSPEX3	(2)(3)	INTTUHCR	6°	0°: 35 ≤6°: 20	-	-	4.4
Interna Universal 		INPSPU0	(3)	INTTUH	15°	0°: 35 ≤15°: 20	INTTUHA 30°	20	4.8
		INPSPU1	(2)(3)	INTTUH	15°	0°: 35 ≤15°: 20	INTTUHA 24°	20	4.8
		INPSPU2	(2)(3)	INTTUH	14°	0°: 35 ≤14°: 20	-	-	4.8
		INPSPU3	(2)(3)	INTTUH	10°	0°: 35 ≤10°: 20	-	-	4.8
		INPSPURO	(1)(3)	INTTUH	15°	0°: 35 ≤15°: 20	INTTUHA 30°	20	4.4

- (1) Rotating interface.
- (2) The final number indicates the gingival height of the interface (measured in mm).
- (3) Indicated for areas of high occlusal load.
- (4) Indicated for areas of low occlusal load.

DIRECT TO IMPLANT									
CONNECTION	Square abutments			Screw included	MAXIMUM ANGULATION screw included	TORQUE RECOMMENDATION screw included	ANGLED SCREW and maximum angulation	TORQUE RECOMMENDATION angled screws (Ncm)	PROSTHETIC Ø (mm)
Externa TINY Narrow platform 		PSPT1	(2)(4)	TTTH	15°	0°: 35 ≤15°: 20	TTUHA 30°	20	4.8
		PSPTR1	(1)(2)(4)	TTTH	15°	0°: 35 ≤15°: 20	TTUHA 30°	20	4.8
Externa Universal 		PSPU1	(2)(3)	TTUH	15°	0°: 35 ≤15°: 20	TTUHA 30°	20	4.8
		PSPU2	(2)(3)	TTUH	15°	0°: 35 ≤15°: 20	TTUHA 30°	20	4.8
		PSPUR1	(1)(2)(3)	TTUH	15°	0°: 35 ≤15°: 20	TTUHA 30°	20	4.8
OVER TRANSEPIHELIAL									
UNIT® (BIOBLOCK® Concept)	Square Interfaces			Screw included	MAXIMUM ANGULATION screw included	TORQUE RECOMMENDATION screw included	ANGLED SCREW and maximum angulation	TORQUE RECOMMENDATION angled screws (Ncm)	PROSTHETIC Ø (mm)
Emergence Ø 3.5 		ISMIUPE	(4)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	3.5
		ISMIUPEX	(3)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	4.5
Emergence Ø 4.1 		ISMIUPU	(3)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	4.1
		ISMIUPUX	(3)	TTMIR	15°	0°: 20 ≤15°: 20	TTMI30 30°	20	5.1
Emergence Ø 5.5 		ISMIUPA	(3)	TTMIUPA	15°	0°: 30 ≤15°: 20	-	-	5.5
		ISMIUPAX	(3)	TTMIUPA	15°	0°: 30 ≤15°: 20	-	-	6.5
Angled/ Emergence Ø 5 		ISMIUPUA	(3)	TTMIR	15°	0°: 30 ≤15°: 20	TTMI30 30°	20	5.5
MULTI-IM® (BIOBLOCK® Concept)	Square Interfaces			Screw included	MAXIMUM ANGULATION screw included	TORQUE RECOMMENDATION screw included	ANGLED SCREW and maximum angulation	TORQUE RECOMMENDATION angled screws (Ncm)	PROSTHETIC Ø (mm)
Emergence Ø 3.5 		ISMIPE	(4)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	4
		ISMIPEX	(3)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	4.5
Emergence Ø 4.1 		ISMIPU	(3)	TTMIR	15°	0°: 20 ≤15°: 20	-	-	4.4
		ISMIPUX	(3)	TTMIR	15°	0°: 20 ≤15°: 20	TTMI30 30°	20	5.1
Emergence Ø 5.5 		ISMIPA	(3)	TTMIR	15°	0°: 20 ≤15°: 20	TTMI30 30°	20	5.5
		ISMIPAX	(3)	TTMIR	15°	0°: 20 ≤15°: 20	TTMI30 30°	20	6.5
Angled / Emergence Ø 5 		ISMIPUA	(3)	TTMIR	15°	0°: 20 ≤15°: 20	TTMI30 30°	20	5.1

- (1) Rotating interface.
- (2) The final number indicates the gingival height of the interface (measured in mm).
- (3) Indicated for areas of high occlusal load.
- (4) Indicated for areas of low occlusal load.