



VISTA APEX

Chlor-XTRA™ (Patented)

REF 503825 - 3mL Pre-filled Syringes

REF 503850 - 12mL Pre-filled Syringes

Description:

Chlor-XTRA™ is a 6% sodium hypochlorite formula; enhanced with powerful wetting agents and proprietary surface modifiers that penetrate into lateral canals and isthmuses by lowering the surface tension, yielding a much more powerful oxidizing agent, 2+ times thinner than standard sodium hypochlorite and 2.6 times more digestive. It is used for the irrigation and debridement of root canals during and after instrumentation.

Instructions for Use:

1. Remove syringe cap and attach irrigating tip securely to syringe.
2. Irrigate with Chlor-XTRA™ as needed for debridement during instrumentation.
3. Remove smear layer with appropriate irrigant, such as EDTA solution, Citric Acid 12.5% or SmearOFF™ 2-in-1.
4. Irrigate again with Chlor-XTRA™ for further debridement.
5. Dispose of syringe and irrigating tip after use.

Precautions:

1. Read MSDS for material handling and safety information.
2. Avoid spilling on surrounding tissues.
3. Avoid spilling on colored fabrics and surfaces as Chlor-XTRA™ is moderately alkaline and has a whitening effect.
4. Proper treatment planning and establishment of working length is required prior to use of this product.
5. Following standard of care, isolate the tooth prior to use of this product to prevent unintended tissue contact or accidental ingestion.

6. Properly evacuate the canal during use of this product to prevent unintended tissue contact or accidental ingestion.
7. It is recommended to use irrigation tips with lateral venting (27ga or smaller) to ensure effective distribution of irrigation solution throughout the root canal and minimize the risk of apical extrusion.

Warnings:

1. Caution, caustic!
2. Keep out of the reach of children.
3. Avoid contact with skin and eyes.
4. Over-instrumentation prior to use of this product can result in unintended tissue contact.
5. Under-instrumentation prior to use of this product can result in tip binding and subsequent apical extrusion. Do NOT bind the tip in the root canal.
6. If tip binding occurs during the procedure, or if resistance is felt during irrigation, discontinue irrigation until the tip binding issue is resolved.
7. NEVER apply excessive pressure when irrigating Chlor-XTRA™.
8. If apical extrusion of Chlor-XTRA™ has occurred, seek medical attention immediately.
9. If ingested, do not induce vomiting. Drink a large portion of water and provide fresh air.

Storage:

- Store under refrigeration.
- Bring solution to room temperature prior to use.

 **DANGER:** Skin corrosion. Serious eye damage. Very toxic to aquatic life. Avoid release to the environment. Wear protective gloves/eye protection/face protection/protective clothing. Harmful if swallowed. Wash thoroughly after handling and avoid any chemical contact with eyes, skin and clothing. Keep container tightly closed to avoid inhalation or accidental ingestion. Use with adequate ventilation when necessary. Follow all government regulations for waste disposal. Discharge to the environment must be avoided. **FIRST AID: IF IN EYES:** Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing. **IF ON SKIN:** Take off contaminated clothing and shoes immediately. Wash off with plenty of soap and water. Consult a physician. **IF INHALED:** Move person to fresh air. **IF SWALLOWED:** Do not induce vomiting. Rinse mouth with water. Specific treatment: Immediately call a Poison Control Center or doctor/physician.



Inter-Med, Inc.
2200 South St.
Racine, WI 53404 USA
Tel: +1-262-636-9755
www.vistaapex.com



2°C
36°F

10°C
50°F



Rx ONLY